


PULSE GENERATOR

PERFORMANCE CHECK

Model: PVH-M1-N

S.N.: 6710

Date: JUNE 29 93



40 dB ATTEN

10 VOLTS / DIV

PRF = 300 KHz

- a) Output signal amplitude:
 ≥ -30 VOLTS
- b) Pulse width:
 ≤ 1 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+135 VOLTS

5 mA

[Signature]