


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR14-2-C-P

S.N.: 6746

Date: SEPT 30 1993



- a) Output signal amplitude: 0 TO +2000 VOLTS
- b) Pulse width:  $R_L \geq 1000 \Omega$   
200 NS TO 5.0 US
- c) Rise time:  $\leq 60 NS$
- d) Fall time:  $\leq 60 NS$
- e) PRF: 0 TO 1 KHz.
- f) Jitter, stability: OK
- g) Prime power:

500 VOLTS/DIV  
 $R_L = 10K.$   
PRF = 1.0 KHz

120/240 VOLTS  
50-60 Hz

