


PULSE GENERATOR
PERFORMANCE CHECK


Model: *MR-61-P5-PN*

S.N.: *6771*

Date: *OCT 14 1993*



P_{out} R_L = 10K
10 µs / DIV



P_{out} R_L = 10K
100 ns / DIV

- a) Output signal amplitude: *0 TO ± 250 VOLTS*
- b) Pulse width: *R_L > 10K*
0.1 µs TO 1.0 ms
- c) Rise time: *≤ 10 ns*
- d) Fall time: *≤ 20 ns*
- e) PRF: *0 TO 1 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *160 / 2000 V*
50-60 Hz

[Signature]