


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-2A-C-P-M*

S.N.: *6789*

Date: *OCT 21 1993*



- a) Output signal amplitude:
0 TO +5.0 AMP
- b) Pulse width:
0 TO 50 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 2 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

≈ 7 AMPS/DIU
PRF = 10 KHz.

120/240 V
50-60 Hz
AM