

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AVO-1-P-DPA1

S.N.:

6806

Date:

NOV 24 '93



a) Output signal amplitude:

10 TO 40 mV

b) Pulse width:

10 NS (FWHM)

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 3 KHz

f) Jitter, stability:

OK

g) Prime power:

+10 VOLTS

300 mA

max

[Signature]

x 20 AMP/DIV
PRF = 3 KHz.