


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM4-3-C-P*

S.N.: *680*

Date: *OCT 18 1993*



- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width: *(TO 50ns)*
- c) Rise time: \leq *600 ps*
- d) Fall time: *NA*
- e) PRF: *0 TO 25 MHz.*
- f) Jitter, stability: *OK*
- g) Prime power:

50 db ATTEN
232 VOLTS/DIV
PRF \approx 12 MHz

120/240 V
50-160 MHz
[Signature]