


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-2-PN-MPS-ET-2W*

S.N.: *6345*

Date: *JUN 24 94*



- a) Output signal amplitude:
0.70 ± 15 VOLT
- b) Pulse width:
0.2 TO 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK

40 dB ATTEN

10 VOLT/DIV

PRF ≈ 10 MHz

- g) Prime power:

120/200 V

50-60 Hz

OK