


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-3-C-P-ATT2*

S.N.: *6862*

Date: *DEC 31 93*



*ME OUT
20 dB ATTEN
100 AMP/DIU*

- a) Output signal amplitude:
f ≤ 100 MHz: 0 to 700 mA
f = 155 MHz: 0 to 600 mA
f = 200 MHz: 0 to 400 mA
- b) Pulse width:
300 to 1.5 μs
- c) Rise time:
≤ 150 ps
- d) Fall time:
≤ 150 ps
- e) PRF: *50 to 250 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 - 60 MHz

[Signature]