


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-P*
S.N.: *6866*
Date: *FEB 16 1994*



50 db ATTEN
32 VOLTS/PW
FBF = 100 KHz.

- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS,
150 mA (max)