


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-6C-C-P-1M4A-0P2*


S.N.: *6871*

Date: *APRIL 5 1994*



- a) Output signal amplitude:  
*0 TO + 5.0 AMP*
- b) Pulse width:  
*50 NS TO 5.0 μS*
- c) Rise time:  
*(1% MAX DUTY CYCLE)  
≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF:  
*0 TO 10 KHz.  
(9% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*

- g) Prime power:  
*120/240 V RMS,  
50-60 Hz.*



TOP *I<sub>out</sub> & PW NEAR MAX.  
PRF = 1 KHz.*

BOOT *I<sub>out</sub> NEAR MAX  
PW NEAR MIN  
PRF = 10 KHz*