


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-101-P5-P-2A*

S.N.: *6872*

Date: *MAR 1 1994*



60 dB ATTEN

100 50V/83/10U

PRF = 30 KHz

- a) Output signal amplitude:
0 TO 700 VOLT
- b) Pulse width:
≤ 1.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/200 VOLT
50-60 Hz.