

PULSE GENERATOR
PERFORMANCE CHECK

Model: *ADH-H21-P*

S.N.: *6006*

Date: *FEB 24 1994*



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
6 1.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
200 mA (max)

[Signature]