


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-P*  
 S.N.: *6915*  
 Date: *FEB 7 94*



- a) Output signal amplitude:  
*0 TO +50 VOLTS (TO 50V)*
- b) Pulse width:  
*2 TO 100 NS*
- c) Rise time:  
*5 NS/DIV ≤ 350 ps*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*1 NS/DIV 0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15 VDC*  
*250 mA (MAX)*

*50 db ATTEN*  
*∴ 32 VOLTS/DIV*  
*10 KHz PULF.*  
*TOP* *5 NS/DIV*  
*BOT* *1 NS/DIV*  
*(RISE TIME)*

*[Signature]*