


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-442-C-N-PN

S.N.: 6987

Date: APRIL 21 1994



- a) Output signal amplitude:
0 TO 720 VOLTS
- b) Pulse width:
0.3 TO 2.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

40 db ATTEN
10 VOLTS/PN
PRF = 100 KHz

120 / 240 V
50 - 60 Hz