


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVMP-2-CP*

S.N.: *6990*

Date: *APRIL 3 1994*



40 db ATTN

10 VOLT/ DIV

PRF = 100 kHz

- a) Output signal amplitude:
0 TO +10 VOLT
- b) Pulse width:
5 TO 100 nS
- c) Rise time:
≤ 100 pS
- d) Fall time:
≤ 135 pS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLT
50-60 Hz

[Signature]