


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-3-C-M5*


S.N.: *7019*

Date: *JUN 13 1994*



- a) Output signal amplitude:  
*0 TO +30 VOLTS*
- b) Pulse width:  
*8 NS TO 1.0US*
- c) Rise time:  
*≤ 300 PS*
- d) Fall time:  
*≤ 300 PS*
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*170/240 V*  
*50-60 Hz*

Ⓐ *40 dB ATTEN*  
*10 VOLTS/DIV*  
*PRF = 10 KHz*



Ⓑ *40 dB ATTEN*  
*10 VOLTS/DIV*  
*PRF = 10 KHz*

*[Signature]*