


PULSE GENERATOR  
PERFORMANCE CHECK

Model: 17V4-17V1-P

S.N.: 7037

Date: JUNE 13 1994



60 dB ATTEN

100 V0.43/DIU

PRF = 50 KHz

- a) Output signal amplitude:  
0 TO +100 V0.43
- b) Pulse width:  
≤ 10 NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

415 VDC

200 mA  
(MAX)

*[Signature]*