


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ARR-G1-PS-P-CS1-R5*


S.N.: *7045*

Date: *JUNE 21 1994*



- a) Output signal amplitude:  
*0 TO +300 VOLTS*
- b) Pulse width: ( $R_L \geq 10K$ )  
*100 NS TO 1.0 MS*
- c) Rise time:  
*10 NS TO 200 NS*
- d) Fall time:  
*10 NS TO 200 NS*
- e) PRF: *0 TO 1 KHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*150/200 VOLTS*  
*50-60 Hz*

*(A)  $R_L = 10K$ .  
RISE, FALL TIME SET  
AT MIN.*



*(B) AS (A) BUT RISE TIME  
ONE TURN CONTROL  
FULLY CW.*