


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-421-C-PN-M-05*
 S.N.: *7096*
 Date: *Aug 31 1994*



- a) Output signal amplitude:
0 TO ± 50 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 250 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 Hz
[Signature]

50 dB ATTEN
= 32 VOLTS/DIV
TRF 5 NS/DIV
PROT. 500PS/DIV (RISE TIME)