


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-7D-C-N*


S.N.: *7109*

Date: *NOV 18 1994*



- a) Output signal amplitude:  
*0 TO -100 AMPS*
- b) Pulse width: *(TO 0 TO 100V  
2μS TO 200μS  
(0.1% MAX DUTY CYCLE))*
- c) Rise time:  
*≤ 1μS*
- d) Fall time:  
*≤ 1μS*
- e) PRF: *0 TO 100 Hz  
(0.1% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V  
50-60Hz*

④ *R<sub>L</sub> = 0.5Ω PRF ≈ 100 Hz*



⑤ *R<sub>L</sub> = 0.5Ω PRF ≈ 10 Hz*

*[Signature]*