


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-HV2-P-EA-EW-CEA*

S.N.: *7124*

Date: *SEPT 30 1994*



- a) Output signal amplitude:
0 TO +100.0V
- b) Pulse width:
0 TO 20NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC

60 db ATTEN
100 VOLTS/DIV
PRF = 20 KHz

250 mA (MAX)

AS