


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AL-155A-179MA*

S.N.: *7126*

Date: *SEPT 21 1994*



- a) Output signal amplitude: *0 TO -170 mA*
- b) Pulse width: *20 NS TO  $\infty$*
- c) Rise time:  *$\leq 10 NS$*
- d) Fall time:  *$\leq 10 NS$*
- e) PRF: *DC TO 40 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *-15V DC 200 mA*  
*+5-BVDC 40 mA*

