


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-5-1-PS-P*  
 S.N.: *7127*  
 Date: *JAN 11 1995*



- a) Output signal amplitude:  
*0 TO +10 VOLT*
- b) Pulse width:  
*≤ 130 PS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability:  
*O.K.*
- g) Prime power:  
*120 / 240 V*  
*50-60 Hz*  
~~*NA*~~

*40 db ATTEN.*  
*10 V 283 / DIV*  
*TDR 1 NS / DIV*  
*BOT 100 PS / DIV*