


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-1A02-PE1A-P

S.N.: 7130

Date: OCT 14 1994



- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width: (TO 50 NS)
8 TO 50 NS
- c) Rise time:
 $\leq 2 \text{ NS}$
- d) Fall time:
 $\leq 2 \text{ NS}$
- e) PRF: 0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
375 mA (MAX)

[Signature]

60 db ATTEN
100 VOLTS/DIV

PRF = 10 KHz

$V_{AMP} = 10 \text{ VOLTS}$

$V_{PW} = 10 \text{ VOLTS}$