


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PV-155-PS-HW1*

S.N.: *7133*

Date: *JAN 11 1995*



5 V/DIV, 500 NS/DIV.



*5 V/DIV, 500 NS/DIV
 $R_L \approx 3 \Omega$
(RISE TIME)*

- a) Output signal amplitude:
0 TO ± 2.5 Amms
- b) Pulse width:
1 TO 10 μ S
- c) Rise time:
 ≤ 100 NS
- d) Fall time:
 ≤ 100 NS
- e) PRF:
0 TO 100 Hz.
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz.*

[Signature]