


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVMP-4-C-PN*

S.N.: *7135*

Date: *FEB 7 1995*



*40 dB ATTEN
10 VOLTS/DIV
PRF = 100 KHz*



*R_c = 50 Ω
PRF = 1 KHz
5 VOLTS/DIV*

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- P_{OUT}* b) Pulse width:
*10 NS TO 10 μS
(10% MAX DUTY CYCLE)*
- c) Rise time:
≤ 200 PS
- N_{OUT}* d) Fall time:
≤ 200 PS
- e) PRF: *0 TO 100 MHz
(10% MAX DUTY CYCLE)*
- f) Jitter, stability:
OK
- g) Prime power:
*120 / 240 VOLTS
50 * 60 KHz*

[Signature]