


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-HU2-P-CA-EW-COA*
S.N.: *7139*
Date: *NOV 21 1994*



- a) Output signal amplitude:
0 TO 100 VOLTS
- b) Pulse width:
0 TO 20NS
- c) Rise time:
≤ 2NS
- d) Fall time:
≤ 2NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VDC.
250mA. (MAX)
eb

60 dB ATTEN.
-100 VOLTS DIV
PRF ≈ 10 KHz