


PULSE GENERATOR
PERFORMANCE CHECK

Model: ARR-61-P-C-0543

S.N.: 7149

Date: NOV 1 1994



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: $CR_L \geq 50 \mu$
50 NS TO 1.0 μS
- c) Rise time:
 $\leq 10 \text{ NS}$
- d) Fall time:
 $\leq 10 \text{ NS}$
- e) PRF: 0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 VOLTS
50 - 60 Hz.
- h) DELAY A-B: 0.1 TO
10 μS.

50 V/DIV
0.5 μS/DIV
 $R_L = 1 \text{ MEG}$
PRF = 10 KHz.