


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *M-1010-C*

S.N.: *7168*

Date: *DEC 20 1964*



Ⓐ *T<sub>out</sub> R<sub>L</sub> = 50 Ω*



Ⓑ *N<sub>out</sub> R<sub>L</sub> = 50 Ω*

a) Output signal amplitude:

*0 TO ± 100 VOLTS*

b) Pulse width: *TO R<sub>L</sub> > 50 Ω*

*50 NS TO 10 MS*

c) Rise time: *TO 7.5 MA*

*DUTY CYCLE*

d) Fall time: *≤ 15 NS*

*≤ 15 NS*

e) PRF: *0 TO 1.0 MHz*

f) Jitter, stability: *(TO 7.5 MA DUTY CYCLE)*

*OK*

g) Prime power:

*120/240 V*

*50-60 Hz*

*[Signature]*