


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MH-5-1-e-P*

S.N.: *7175*

Date: *JAN 9 1995*



- a) Output signal amplitude:
0 TO + 10 VOLT
- b) Pulse width:
≤ 130 PS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLT

570 160 Hz
ff

40 dB ATTEN

10 VOLTS/DIV

TOP 2.0 NS/DIV

BOT 200 PS/DIV

PRF ≈ 100 KHz