


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR4 - .3 - PS - PN

S.N.: 7198

Date: JAN 19 1995



- a) Output signal amplitude:
0 TO ± 3000 V.
- b) Pulse width: $R_L \geq 10K$
200 NS TO 2.5 μ S
- c) Rise time:
 ≤ 100 NS
- d) Fall time:
 ≤ 100 NS
- e) PRF: 0 TO 1 KHz.
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

500 VOLTS/DIV
 $R_L = 10K$
PRF = 1.0K

