


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1423-C-P-M*

S.N.: *7203*

Date: *JAN 26 1995*



- a) Output signal amplitude:
0 TO + 40 VOLTS
- b) Pulse width:
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 VOLTS
50 - 60 Hz

50 dB ATTEN

≈ 32 VOLTS/DIV

PRF ≈ 100 KHz

ALL