


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2-C-P*
S.N.: *7206*
Date: *DEC 19 1994*



- a) Output signal amplitude:
0 TO +10 AMP(S)
- b) Pulse width:
0 TO 15 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 1 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 7 AMP(S) / DIV
PRF = 20 KHz

120/240 VOLT
50-60 Hz