


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-155C-C-MA-P-P-11  
S.N.: 7230  
Date: MAR 27 1995



- a) Output signal amplitude:  
PULSE: 0 TO +2.0 AMPS  
DC: 0 TO 500
- b) Pulse width:  
160 NS TO 10 US
- c) Rise time:  
 $\leq 50$  NS
- d) Fall time:  
 $\leq 50$  NS
- e) PRF:  
0 TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

120 / 240 V  
50-60 Hz

2 VOLTS / DIV  
 $R_L = 1.0 \Omega$   
 $\approx 2$  AMPS OUT (PULSE)  
DC  $\approx 0$ .