


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ARR-AA-PW-C-PN*


S.N.: *7251*

Date: *APRIL 20 1995*



- a) Output signal amplitude:
0 TO ± 400 VOLTS
- b) Pulse width: *R_L > 50Ω*
50NS TO 5.0μs
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
≤ 20NS
- d) Fall time:
≤ 20NS
- e) PRF: *0 TO 10 KHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: *(CYCLE)*
OK
- g) Prime power:
120/240 V
50-60Hz

(A) P_{out}: R_L = 50Ω PRF = 1 KHz



*(B) P_{out}: R_L = 50Ω
PRF = 5 KHz*

*(CHECKED FOR
90 VOLT
50-60Hz)*