


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PX-CP-1-021*

S.N.: *7257*

Date: *MAR 15 1995*



- a) Output signal amplitude:
PW_m = 2 ns, ≈ 14 V_{0.8}
- b) Pulse width:
PW_m = 0.2 ns, ≈ 10 V_{0.8}
0.2 TO 2.0 ns
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 400 ps
- e) PRF: *0 TO 25 MHz*
- f) Jitter, stability:
OK
- g) Prime power:
-1.5 TO -2.5 V
(2 mA max)

40 & 6 ATTEN
10 V 283 / 10V
INPUT AMP ≈ 15 V_{0.8}