


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5B.1-C-P-7E02*

S.N.: *7260*

Date: *MAY 10 1995*



- a) Output signal amplitude:  
*0 TO + 75 AMPS*
- b) Pulse width: *(TO 0.6μ)*
- c) Rise time: *50 TO 500 NS*  
*≤ 15 NS*
- d) Fall time:  
*≤ 15 NS*
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 VOLTS*  
*50-60 HZ*

*AM I OUT TO 50Ω*  
*≈ 2 VOLTS / DIV.*

*120/240 VOLTS*  
*50-60 HZ*  
*[Signature]*