


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-MPT-P*

S.N.: *7266*

Date: *APRIL 18 1995*



a) Output signal amplitude:

≈ 30 VOLT

b) Pulse width:

≤ 1 ns

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 to 1 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 VDC

120 mA (MAX)

40 dB ATTEN

10 VOLT/DIV

PRF ≈ 300 kHz

[Signature]