


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-1N3-P-PS

S.N.: 7304

Date: JUNE 12 1995



- a) Output signal amplitude:
0 TO +60 VOLTS
- b) Pulse width:
≤ 1.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:

50 dB ATTEN

= 32 VOLTS/DIV

PRF = 100 KHz.

120/240 V
50-60 Hz