


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVNC-2-C

S.N.: 7318

Date: JUNE 22 1995



- a) Output signal amplitude:
ECL LEVELS
-0.9 + -1.2V (TO -50 Ω)
- b) Pulse width:
TO -2V
- c) Rise time:
50% DUTY CYCLE
 ≤ 1 NS
- d) Fall time:
 ≤ 1 NS
- e) PRF: 10 KHz TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

TOP: RANGE A
 ≈ 50 MHz
5 ns/DIV

120 / 240 VOLTS
30 - 60 MHz
A

BOI RANGE B
 ≈ 250 MHz
2 ns/DIV

