


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-C-PN-05-SP-DNA1-0P2*

S.N.: *7324*

Date: *AUG 9 1995*



- a) Output signal amplitude:  
*0 TO  $\pm 350$  V*
- b) Pulse width: (*50 ns*)  
*5 TO 100 ns*
- c) Rise time:  
 *$\leq 2$  ns*
- d) Fall time:  
 *$\leq 3$  ns*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*

*70 db ATTEN*  
 *$\approx 320$  VOLTS/DIV*  
*PRF  $\approx 2.5$  KHz*

*WY*