


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVL-2-C-PN-05-SP-DNAH-OP2*

S.N.: *7325*

Date: *AUG 9 1995*



- a) Output signal amplitude:  
*0 TO  $\pm$  350 VOLTS*
- b) Pulse width: *(50 ns)*
- c) Rise time: *5 TO 100 ns*  
 *$\leq$  2 ns*
- d) Fall time:  *$\leq$  3 ns*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*

*70 db ATTEN*  
 *$\approx$  320 VOLTS/DIV*  
*PRF = 2.5 KHz*

*MB*