


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH - HVI - PN - SSC1

S.N.: 7333

Date: JULY 31, 1995



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+ 15 VOLTS
250 MA MAX

TOP: P out
BOT: N out
60 db ATTN
100 VOLTS / DIV
PRF = 30 KHz