


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-2-C-OT-AM-P*

S.N.: *7356*

Date: *JUN 11 1995*



- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 VOLTS*

*TOP: 5 NS/DIV*  
*TRIM: 200 PS/DIV*  
*(RISE TIME)*  
*FOT: 200 PS/DIV*  
*(FALL TIME)*  
*40 dB ATTEN*  
*10 VOLTS/DIV*  
*0.5 MHz*