


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-6.1-C-P-CS1-R5*


S.N.: *7398*

Date: *AUG 30 1995*



- a) Output signal amplitude:  
*0 TO + 300 VOLTS*
- b) Pulse width: *R<sub>L</sub> > 10K*  
*100NS TO 1MS*
- c) Rise time:  
*10NS TO 300NS*
- d) Fall time:  
*10 TO 300NS*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60Hz*

Ⓐ *20ns/DIV, 50V/DIV*  
*R<sub>L</sub> = 33K.*



Ⓑ *100 NS/DIV, 50V/DIV*  
*RISE TIME AT MIN, MID*  
*& MAX VALUES.*

*[Signature]*