


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-A-1-52-P-C-HT2-6A*

S.N.: *7412.*

Date: *NOV 6 1995*



- a) Output signal amplitude: *0 TO +200 VOLTS*
- b) Pulse width: *( $R_L = 50 \Omega$ )*
- c) Rise time: *10 TO 100 NS*  
 *$\leq 2 NS$*
- d) Fall time:  *$\leq 2 NS$*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *100/200 VOLTS*  
*( $\pm 10\%$ )*

*60 dB ATTEN*

*= 100 VOLTS / DIV*

*PRF = 100 KHz.*

*100/200 VOLTS*  
*( $\pm 10\%$ )*

*50 - 60 kHz.*