


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-61-C-P*

S.N.: *7422*

Date: *JAN 12 1996*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
*($R_L \approx 100K$)
0.545 TO 0.5 μ S*
- c) Rise time:
*(CONTROLLED BY PRF)
 $\leq 10NS$*
- d) Fall time:
 $\leq 10NS$
- e) PRF:
GND 1 KHZ TO 1 MHZ
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 VOLTS
50-60 HZ*

*$R_L = 100K$
PRF ≈ 1 MHZ*