


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AV-141C-PS

S.N.: 7437

Date: SEPT 14 1995



a) Output signal amplitude:

3 VOLT P-P

b) ~~Pulse width:~~

GAIN: 20 dB

c) Rise time:

5 NS

d) Fall time:

5 NS

e) PRF:

DC TO 600 MHz

f) Jitter, stability:

OK

g) Prime power:

120/240 V

50-60 Hz

*[Signature]*

49 dB ATTEN

5 VOLT/DIV

5 NS/DIV