


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AV-1-0A-ECL-N-M*

S.N.: *7443*

Date: *OCT 3 1995*



- a) Output signal amplitude:
0 TO -10 VOLTS.
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 100 ps
- d) Fall time:
≤ 135 ps
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS
400 mA (max)

(A)

40 db ATTEN
-10 VOLTS/DIV
TDP: 1 NS/DIV
DOT: 200 PS/DIV