


PULSE GENERATOR

PERFORMANCE CHECK

Model: AMH-5-1702-PVI-C-P*

S.N.: 7447

Date: OCT 6 1995



- a) Output signal amplitude: 0 TO +30 VOLTS.
- b) Pulse width: ≤ 350 PS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 2 MHz*
- f) Jitter, stability: OK
- g) Prime power: 120/240 V
50-60 Hz

40 dB ATTEN
10 VOLTS/DIV
PRF = 2.0 MHz

* -PVI OPTION,
PRF MAX IS
2 MHz RATHER
THAN 1 MHz